



Centrum für Hygiene
und medizinische Produktsicherheit

Certificate

for the company

Kerox Kft.
H-2049 Diósd
Homokbánya út 77

manufactured


zircon blocks for the dental technology

can be certified after testing
Cytotoxicity according DIN EN ISO 10993-5, 2009
testing for irritation and delayed-type hypersensitivity according to the
DIN EN ISO 10993-10, 2009 ,
Testing of genotoxicity according ISO 10993-3, 2009


the biological harmlessness according DIN EN ISO 10993-1 and ISO 7405.

Identification Number: 99198

Schwerin, 09/06/2010


Dipl.-Ing (FH) Margrit Werner
Executive Manager




Prof. Dr. med. H.-P. Werner
Scientific-Technical Director